



Certificate of Compliance

Certificate: 2538661

Master Contract: 157918

Project: 2731406

Date Issued: October 15, 2014

Issued to: BW Technologies by Honeywell

2840 2nd Ave SE
Calgary, AB T2A 7X9
Canada
Attention: Kevin Giddings

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Alan McLaughlin

Issued by: Alan McLaughlin, P.Eng.

PRODUCTS

CLASS 2252 06 - PROCESS CONTROL EQUIPMENT

CLASS 2252 86 - PROCESS CONTROL EQUIPMENT - Certified to US Standards

Portable gas detector automatic calibration system, Model IntelliDoX Docking Module, rated 12V DC, 8.33 A 5 modules max ganged.

Notes:

- The above model is powered by an external Certified Power Supply, Equipment Class III, Pollution Degree 2.
- Mode of operation: Continuous
- Environmental Conditions: Normal: 5 to 40 C, 2000 m max, 80% to temperatures up to 31 °C decreasing linearly to 50% rH at 40 °C;

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

UL Std. No. 61010-1 (3rd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements